



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Gillmore Needle Apparatus ZI 1006

Standards:ASTM C 266

Specification:

A base with a Vertical shaft and Two horizontal arms. The lower arms is adjustable for height. 1 no. Initial needle 1/12-inch dia ¼ lb. Wt. 1 no.Final needle 1/24 inch dia. ¼ lb. Wt. 1 no. Glass base plate.Complete as above.

