



**ZEAL INTERNATIONAL** (An ISO 9001 Certified Company)

**Head Office:** 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

**Branch Office:** 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

**Warehouse:** 4/26 Kirti Nagar Industrial Area New Delhi – 110015

[info@zealinternational.com](mailto:info@zealinternational.com)

[sales@zealinternational.com](mailto:sales@zealinternational.com)

[support@zealinternational.com](mailto:support@zealinternational.com)

[quotation@zealinternational.com](mailto:quotation@zealinternational.com)

Visit us at: [www.zealinternational.com](http://www.zealinternational.com)

## Le Chatelier Flask ZI 1012

Standards: IS 4031 1968, ASTM C 188

Specification:

Used for finding specific gravity of hydraulic cement. Made from Borosilicate glass. The flask is 243mm in total height, having a bulb of 90mm dia of 250ml approximate capacity. The long neck of the flask has at top a funnel of 50mm dia in that fits a ground glass stopper. The neck has over-all 11mm i.d. upper portion is graduated from 18ml to 24ml with 1 ml graduation. Just at the bottom of the neck 1 ml capacity is marked in between there is 17 ml capacity bulb.

