



**ZEAL INTERNATIONAL** (An ISO 9001 Certified Company)

**Head Office:** 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

**Branch Office:** 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

**Warehouse:** 4/26 Kirti Nagar Industrial Area New Delhi – 110015

[info@zealinternational.com](mailto:info@zealinternational.com)

[sales@zealinternational.com](mailto:sales@zealinternational.com)

[support@zealinternational.com](mailto:support@zealinternational.com)

[quotation@zealinternational.com](mailto:quotation@zealinternational.com)

Visit us at: [www.zealinternational.com](http://www.zealinternational.com)

## Heat of Hydration Apparatus ZI 1017

Standards: IS 11262-1985, ASTM C 186

Specification:

This equipment is required to determine the heat of hydration of cement as expressed in calories per gram.

The equipment comprises of the following:

1. A wide mounted double walled vacuum flask with a stop cock 38 mm & a insulating container for the flask
2. A Beckman thermometer (Range 5° C) held tightly by the cock stopper in such a way as to avoid accidental contact with the stirrer blade & the reading lens. To facilitate the easy removal the cock stopper is in two halves.
3. A constant speed stirrer (double bladed propeller type) extended to within 38 mm from the bottom of the flask.
4. A funnel (Gooch type) with a stem of 6 mm inner dia & a body approx 25 mm long and 25 mm dia. is fitted to the cock stopper for introducing the sample.

