



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Slake Reactivity Apparatus ZI 1029

Standards:

Specification:

HSN Code: 9024

It is used for checking lime reaction in water under controlled conditions. Unit is supplied complete with stirrer with geared motor, a stainless steel container duly insulated, a teflon sealing cap to protect heat loss from the vessel, a digital temperature indicator with stirring control knob and a stainless steel temperature sensor, plastic measuring cylinder 1ltr, plastic measuring jug 1ltr, digital stop clock and digital portable pen type thermometer. Unit is mounted on a heavy duty stand for easy operation.

