



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Skid Resistance And Friction Tester Standard ZI 5017

Standards:

Specification:

EN 13036-4 (Also Conforming to EN 1097-8, ASTM E103)

Used for the measurement of surface friction properties, the apparatus is suitable for both site and laboratory applications and for polished stone value tests using curved specimens from accelerated polishing tests. The equipment is supplied with:

- Additional scale for tests on polished stone value specimens
- 6 rubber sliders for site use, complete with conformity certificate
- Thermometer -10 to +110°C for surface temperature measurement
- 1 litre washing bottle, for surface wetting
- Tool set with case, for machine assembly
- Rule for sliding length verification

Specification:

- Rocker weight : 1500 +/- 30 g
- Distance between rocking centre & center of gravity : 410 +/- 5 mm
- Positive static pressure on pavement : 22.2 +/- 0.5 kn
- 6 rubber slider and base plate dimension (mm) : 700x360x70

