



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Hardness Tester For Mastic Asphalt ZI 4011

Standards:IS 1195

Specification:

For determining the hardness number of Mastic Asphalt for flooring.

It consist of an internally insulated cabinet to the base of which is fitted a water bath having two taps. The bath is heated by an immersion heater and the temperature is controlled with the thermostat at $35^{\circ} \pm 0.5^{\circ} \text{C}$. A 6.5mm dia pin is loaded on the specimen by means of a lever to give a 31.7 Kg weight. A dial gauge 0.01 x 25 mm is provided to record the penetration of pin into the specimen.

