



**ZEAL INTERNATIONAL** (An ISO 9001 Certified Company)

**Head Office:** 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

**Branch Office:** 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

**Warehouse:** 4/26 Kirti Nagar Industrial Area New Delhi – 110015

[info@zealinternational.com](mailto:info@zealinternational.com)

[sales@zealinternational.com](mailto:sales@zealinternational.com)

[support@zealinternational.com](mailto:support@zealinternational.com)

[quotation@zealinternational.com](mailto:quotation@zealinternational.com)

Visit us at: [www.zealinternational.com](http://www.zealinternational.com)

## Centrifuge Extractor (Hand Operated) ZI 4013

Standards: ASTM D2172, AASHTO T-58, T-164

Specification:

The Instrument is used for determination and checking of Bitumen percentage in Bituminous mix, the mix is added with a solvent and dissolved bitumen is removed by centrifugal action. Consists of a removable Aluminum rotor bowl, Capacity 1500 gms.

With a cap and tightening nut. The bowl assembly is mounted on a vertical shaft, which protrudes from a cast housing. This shaft and thus the bowl is rotated fast manually by enclosed gears in the cast body and handle. Solvent is introduced during the test through the holes in the cap of the housing. A drain is provided to collect dissolved Bitumen coming out of the rotating bowl and getting collected in the housing.

