



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Centrifuge Extractor (Motorised) (ASTM) ZI 4015

Standards:ASTM: D2172 AASHTO: T-58, T-164

Specification:

Used for the determination of bitumen percentage in bituminous mixtures. It consists of a removable, precision machined aluminium rotor bowl (accessory 1500 or 3000 g capacity), housed in a cylindrical aluminium box. The separate control panel incorporates an electronic card fitted with AC drive that automatically drives the bowl speed rotation ramp from 0 to 3600 R.P.M. as requested by Standards, with automatic fast stop bowl rotation at the end of the test. Supplied complete with speed regulator and digital display monitoring the frequency. Power supply: 230 V A.C. Single Phase

