



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Universal Water Bath (Serological Type) ZI 9011

Standards:

Specification:

Double walled inner chamber in stainless steel outer chamber in mild steel temperature range Ambient to 100° C controlled by the capillary thermostat. Accuracy $\pm 2^{\circ}$ C operated on 230V A.C. Single Phase. These baths are suitable for washing Marshall specimen. The lid is made up of stainless steel with a hole for glass thermometer to penetrate.

Following sizes are available:

1. 12"lengthx 9" width x 7" height having rating 750 W.
2. 18"length x 12" width x 7" height having rating 1500 W.

Customised sizes also available. Rack also available for Le Chatelier mould for 12 Nos.

