



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

pH Tester ZI 9048

Standards:

Specification:

ZI 9048 feature a rugged and ergonomically enhanced casing that fits comfortably in your hand and a larger LCD which makes it easier to read at any angle. With the latest innovation of a non-clogging fiber junction ZI 9048 meters last much longer. Simply pull out 1-2 mm (1/8") of the junction fiber to completely renew the junction.

ZI 9048 offers a Range of 0.0 to 14.0 pH and an Accuracy of ± 0.1 pH.

This tester is calibrated manually to 2 points.

Specifications:

- pH Range : 0.0 to 14.0 pH
- pH Resolution : 0.1 pH
- pH Accuracy : ± 0.1 pH
- Calibration : manual, 2 points
- Battery Type / Life : 4 x 1.5V / approx. 700 hours of continuous use
- Environment : 0 to 50°C (32 to 122°F); RH max 95%
- Dimensions : 175 x 41 x 23 mm (6.9 x 1.6 x 0.9")
- Weight : 95 g (3.4 oz.)

