



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Slump Cone, Plastic ZI 2003

Standards: BS 1881:102, ASTM C143, EN 12350-2, EN 12350-8

Specification:

Slump Cone test is performed to determine the consistency as well as workability of fresh concrete.



Slump Cone, Plastic, supplied complete with:

1. Slump Cone, plastic. Weight: 750 gm
2. Tamping Rod, steel, 16 mm (dia.) x 600 mm (length)
3. Base Plate, galvanized steel
4. Stainless Steel measuring ruler, 300 mm (length)
5. Aluminium Scoop, capacity: 500 cc
6. Fine Wire Brush
7. Weight: approx. 5 kg