



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Kelley Ball Penetration Apparatus ZI 2006

Standards: ASTM C-360

Specification:

The apparatus is used to determine the workability of Portland cement & concrete. The Kelly ball test is considered to be simple and much faster than the slump test. Twice the Kelly ball reading approximately equals the slump. It consists of a cylinder with a ball shaped bottom and handle, together weighing 15 kg. A strip frame, guides the handle and serves as a reference for measuring the depth of penetration. The handle is graduated in mm. Penetration can be recorded to the nearest 0.5 mm.

