



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Concrete Flow Table Test Set ZI 2010

Standards: EN 12350-5

Specification:



The test set is used for concrete mixes of high workability and determines flow index as an arithmetic mean of the diameter of the specimen after working on a flow table. The apparatus consists of a double steel table, an upper table measuring 700 x 700 mm and hinged at one side to the lower table. The top table is inscribed and all parts are protected against corrosion. The stainless steel cone has a 130 ±2 mm top diameter, 200 ±2 mm base diameter and 200 ±2 mm height and 1.5 mm thickness. Dimensions: 700 x 850 x 300 mm, Weight (approx.) 40 kg

ZI 2010 : Concrete flow table test set

ZI 2010-A : Flow cone for

ZI 2010-B : Wooden tamper 40 x 40 x 335 mm