



**ZEAL INTERNATIONAL** (An ISO 9001 Certified Company)

**Head Office:** 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

**Branch Office:** 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

**Warehouse:** 4/26 Kirti Nagar Industrial Area New Delhi – 110015

[info@zealinternational.com](mailto:info@zealinternational.com)

[sales@zealinternational.com](mailto:sales@zealinternational.com)

[support@zealinternational.com](mailto:support@zealinternational.com)

[quotation@zealinternational.com](mailto:quotation@zealinternational.com)

Visit us at: [www.zealinternational.com](http://www.zealinternational.com)

## Concrete Test Hammer (Schmidt Hammer) N Type ZI 2068

Standards: EN 12504-2, 13791, ASTM C 805, BS 1881:202, NF P18-417, DIN 1048, UNI 9189

Specification:



The quality of concrete is mainly judged by its compressive strength directly affecting the loadbearing capacity and durability of concrete structures. ZI 2068 concrete test hammer (Schmidt Hammer N type) is used to measure the compressive strength characteristics of hardened concrete non-destructively, control uniform concrete quality and detect weak spots in the concrete. The test object should have a minimum thickness of 100 mm.