



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Methylene Blue Test Set ZI 3013

Standards: EN 933-9; NF P94-068; UNE 83 180;
UNI 8520-15

Specification:

Set consists of

- High speed agitator motor, 400/600 r.p.m.
- Stirring propeller, \varnothing 70 mm 4 flanks
- Glass burette, 50 ml x 0.1 ml
- Burette holder and stand
- Filter paper, 1 pack (100 pcs.), 125 mm dia, 95 g/m², 0.20 mm thickness
- Glass rod, \varnothing 8x300 mm
- Plastic beaker, 1000 ml
- Methylene blue, 100 g
- Kaolinite, 500 g

