



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Permeability Apparatus (Falling Head Permeability) ZI 3030

Standards: IS 2720 (PART XVII) 1966

Specification:

The apparatus is used for the laboratory determination of permeability of soil using a constant or a variable head. This test is recommended for soils with coefficient permeability in the range of 0.10^{-3} to 10^{-7} cm/sec. The maximum particle size of the soil which can be tested in the mould is 10mm

