



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Universal Permeability ZI 3031

Standards:IS 2720 (PART XVII)

Specification:

Same as ZI 3030 but instead of a set of three glass stand pipes a stand with nine glass tubes of 6mm, 10mm, 20mm, 25mm, 40mm, 50mm, 60mm, 70mm, 75mm, bore tube is provided with over flow arrangement for constant head tests. The remaining tubes are used for falling head test.

