



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Consolidation Apparatus (Single Gang) ZI 3049

Standards: IS 2720 (PART XV)

Specification:

The standard outfit comprises of the following items. Loading unit, maximum capacity 20kg/cm². Having a loading yoke connected to a lever arm with a counter balancing adjustment and having a lever ratio of 1:1 the whole assembly being mounted on a sturdy steel frame stand. The loading unit is so designed that it can be used for consolidation cells of different diameters as well as different dia floating ring type consolidation cells. Fixed ring type Consolidometer (Odeometer) cell assembly for testing 60mm dia x 20mm thick specimen comprising: Fixed ring for specimens 60mm dia x 20mm thick with guide ring. Top and bottom porous stones for 60mm dia specimen. Perforated pressure pad, channeled base with water inlet and gasket Flanged water jacket, water reservoir with plastic tube and pinch cock. Set of weights to give a pressure of 10kg/cm.sq. On 60mm dia specimen, comprising : 7 nos. 0.05 kg/cm², 5nos. 0.1 kg/cm², 6nos. 0.5kg/cm², and 5 nos. 1.0 kg/cm².

Supplied complete as above but without dial gauge.

Accessories :

Dial gauge 0.002 mm x 5mm. Extension piece, 40 mm long, consolidation test forms pad of 50 for one dimensional consolidation.

