



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

AC Digital Clamp Meter JPI 2015

Standards:

Specification:

DC Current 400A, 1200A

DC Voltage 400mV, 4V, 40V, 400V, 600V (Auto)

Frequency 100Hz, 1kHz, 10kHz, 100kHz, 500kHz (Auto)

Sp. Functions Audible Continuity, Diode Test, Capacitance, Peak Hold, Data Hold, Max, APO, Δ ZERO

AC Current 400A, 1000A (TRMS)

AC Voltage 4V, 40V, 400V, 600V (Auto)

Resistance 400 Ω , 4k Ω , 40k Ω , 400k Ω , 4000k Ω , 40M Ω (Auto)

Capacitance 4nF, 40nF, 400nF, 4 μ F, 40 μ F(Auto)

Jaw Opening 55 mm max.

