



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Lux Meter JPI 2025

Standards:

Specification:

Display	3 1/2 digits 18mm LCD
Measuring Range	1Lux~50000Lux
Over-input	indication of "1"
Accuracy	±4% ±10digits(±5% ±10digits as over 10,000Lux)
Temperature Characteristic	: ±0.1%°C
Sampling Time	0.5second
Repeatability	: ±2%
Dimensions	: 118×70×29mm

