



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Digital Conductivity/TDS Meter JPI 9019

Standards:

Specification:

Functions

Display

Conductivity, Temp. (optional)

3 1/2 digit LED with overload indication

Power Supply

230 V + 10% 50 Hz. AC

Conductivity

Range

0.01 uS/cm to 200mS/cm

Resolution

0.01 uS/cm on the lowest range

Accuracy

+ 0.5% of range + 1 digit

Cell const. adj.

Digital 0.1 to 2.00

Temp. Compensation

0 to 500C manual (automatic with Pt100 optional)

Temp. co-efficient

2%

Temperature (Optional

Model G-3001AT)

Range

0 to 100°C

Resolution

0.1°C

Accuracy

+ 0.5% of range

Sensor

PT100 probe

Accessories

Conductivity Cell Instruction manual
Dust Cover Temp. probe optional

