



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Dashing Moisture Testing Machine JPI 7001

Standards:

Specification:

Principle:- Resistance Measurement

Measuring Range:- 3.5% to maximum over 40%

Accuracy :- $\pm 0.2\%$

Temperature Compensation :- Automatic with PT-100 temperature sensor.

Error conditions:- Under & Over Range etc. Displayed

Display:- 3 digits LED display.

Size (L x B x H in cm) :- 45x25x33 cm (approx.)

Weight :- 24 Kg. (Approx)

Power:- 9V DC Adaptor for 220 volt AC mains or 6(R-20 type) dry cells. (supplied without dry cells)

