



**ZEAL INTERNATIONAL** (An ISO 9001 Certified Company)

**Head Office:** 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

**Branch Office:** 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

**Warehouse:** 4/26 Kirti Nagar Industrial Area New Delhi – 110015

[info@zealinternational.com](mailto:info@zealinternational.com)

[sales@zealinternational.com](mailto:sales@zealinternational.com)

[support@zealinternational.com](mailto:support@zealinternational.com)

[quotation@zealinternational.com](mailto:quotation@zealinternational.com)

Visit us at: [www.zealinternational.com](http://www.zealinternational.com)

## Dashing Moisture Testing Machine JPI 7001

Standards:

Specification:

**Principle:-** Resistance Measurement

**Measuring Range:-** 3.5% to maximum over 40%

**Accuracy :-**  $\pm 0.2\%$

**Temperature Compensation :-** Automatic with PT-100 temperature sensor.

**Error conditions:-** Under & Over Range etc. Displayed

**Display:-** 3 digits LED display.

**Size (L x B x H in cm) :-** 45x25x33 cm (approx.)

**Weight :-** 24 Kg. (Approx)

**Power:-** 9V DC Adaptor for 220 volt AC mains or 6(R-20 type) dry cells. (supplied without dry cells)

