



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Jar Test Apparatus JPI 7060

Standards:

Specification:

This unit is designed for formation test in water Treatment plants. Sample is easily viewed by Diffused Base Illumination with Florescent Tube placed below plastic & black background. Fitted with DC motor of 1/20 HP with variable speed up to 200 RPM. Control by Speed Regulator provided with stainless Steel paddles with adjustable spacers to adjust the Depth of Paddles. To work on 220/230 volts AC.

JPI 8060



Jar Test Apparatus