



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Differential Pressure Gauge JPI 12012

Standards:

Specification:

- two colours dial with indication of the pressure droops filed beyond the 75% of the full scale
- dial graduated in mbar and p.s.i.d.
- max working pressure up to 100 bar
- suitable for being installed outside with a -20°C to +60°C room temperature
- accuracy $\pm 20\%$ of the full-scale
- maximum index hand wich may be zeroed manually from the outside
- any calibration is not required
- the structure is able to bear full inlet pressure by one side of the membrane

