



**ZEAL INTERNATIONAL** (An ISO 9001 Certified Company)

**Head Office:** 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

**Branch Office:** 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

**Warehouse:** 4/26 Kirti Nagar Industrial Area New Delhi – 110015

[info@zealinternational.com](mailto:info@zealinternational.com)

[sales@zealinternational.com](mailto:sales@zealinternational.com)

[support@zealinternational.com](mailto:support@zealinternational.com)

[quotation@zealinternational.com](mailto:quotation@zealinternational.com)

Visit us at: [www.zealinternational.com](http://www.zealinternational.com)

## Flash Point (Closed) Pensky-Martens Apparatus ZI 11005

Standards:

Specification:

For determining the flash point of fuels oil and lubricating oil, bitumen other than cutback bitumen and suspension of solids in liquids, having a flash point above 49°C.

