



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Saybolt Viscometer ZI 11007

Standards:

Specification:

Say bolt Viscometer, Electrically Heated, ASTM D88, D244, AASHTO T72 for the empirical measurement of Say bolt Viscosity of petroleum products at specified temperatures between 70° F and 210° F. This is also used for determining the Saybolt Furol Viscosity of bituminous materials at temperatures of 250°, 275°, 300°, 350°, 400° and 450° F. It comprises one each of cylindrical Oil cup, Universal Tip, Furol Tip, Bath Fitted with immersion Heater mounted on a stand. Dimmer stat for temperature control, Stirrer with shield.

