



**ZEAL INTERNATIONAL** (An ISO 9001 Certified Company)

**Head Office:** 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

**Branch Office:** 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

**Warehouse:** 4/26 Kirti Nagar Industrial Area New Delhi – 110015

[info@zealinternational.com](mailto:info@zealinternational.com)

[sales@zealinternational.com](mailto:sales@zealinternational.com)

[support@zealinternational.com](mailto:support@zealinternational.com)

[quotation@zealinternational.com](mailto:quotation@zealinternational.com)

Visit us at: [www.zealinternational.com](http://www.zealinternational.com)

## Engler Viscometer ZI 11009

Standards:

Specification:

Standard water value:  $51 \pm 1$  s (It is at standard water temperature  $20 \pm 0.1^\circ$ , and 200 ml water; It can be converted into a value at  $25 \pm 0.1^\circ$ , and 50 ml water)

Temperature range:  $0 \sim 100^\circ$

Temperature controlling accuracy:  $\pm 0.1^\circ$

Thermometer: As per T0622; . Graduated flask: 100 ml;

Inner container: Made by stainless steel;

Power of electric heater: 550 W/220V;

Precision of electric stopwatch: 1/100 s

