



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Engler Viscometer ZI 11009

Standards:

Specification:

Standard water value: 51 ± 1 s (It is at standard water temperature $20 \pm 0.1^\circ$, and 200 ml water; It can be converted into a value at $25 \pm 0.1^\circ$, and 50 ml water)

Temperature range: $0 \sim 100^\circ$

Temperature controlling accuracy: $\pm 0.1^\circ$

Thermometer: As per T0622; . Graduated flask: 100 ml;

Inner container: Made by stainless steel;

Power of electric heater: 550 W/220V;

Precision of electric stopwatch: 1/100 s

